



FLEXdose SELECT

Activity Chart

(SIR-Y002 3 mL vial)



SIR-Spheres Y-90 resin microspheres (SIR-Y002 3 mL vial)

DAY FIVE PRE-CALIBRATION

	Elapse Time (in hours)	Decay Correction Factor	SIR-Y002 Activity in GBq	SIR-Y002 Activity in GBq per .5mL	SIR-Y002 Activity in mCi	SIR-Y002 Activity in mCi per .5mL	SIR-Y002 Specific Activity Bq Per Sphere^	SIR-Y002 SIR-Spheres per GBq (reported in millions^)
12:00 AM	-138.0	4.447	8.00	1.33	216.35	36.06	293.22	3.41
1:00 AM	-137.0	4.399	7.92	1.32	214.02	35.67	290.07	3.45
2:00 AM	-136.0	4.352	7.83	1.31	211.72	35.29	286.95	3.48
3:00 AM	-135.0	4.305	7.75	1.29	209.44	34.91	283.86	3.52
4:00 AM	-134.0	4.259	7.67	1.28	207.19	34.53	280.81	3.56
5:00 AM	-133.0	4.213	7.58	1.26	204.96	34.16	277.79	3.60
6:00 AM	-132.0	4.168	7.50	1.25	202.76	33.79	274.80	3.64
7:00 AM	-131.0	4.123	7.42	1.24	200.58	33.43	271.84	3.68
8:00 AM	-130.0	4.079	7.34	1.22	198.42	33.07	268.92	3.72
9:00 AM	-129.0	4.035	7.26	1.21	196.29	32.71	266.03	3.76
10:00 AM	-128.0	3.991	7.18	1.20	194.17	32.36	263.17	3.80
11:00 AM	-127.0	3.948	7.11	1.18	192.09	32.01	260.34	3.84
12:00 PM	-126.0	3.906	7.03	1.17	190.02	31.67	257.54	3.88
1:00 PM	-125.0	3.864	6.96	1.16	187.98	31.33	254.77	3.93
2:00 PM	-124.0	3.822	6.88	1.15	185.95	30.99	252.03	3.97
3:00 PM	-123.0	3.781	6.81	1.13	183.95	30.66	249.32	4.01
4:00 PM	-122.0	3.741	6.73	1.12	181.98	30.33	246.63	4.05
5:00 PM	-121.0	3.700	6.66	1.11	180.02	30.00	243.98	4.10
6:00 PM	-120.0	3.661	6.59	1.10	178.08	29.68	241.36	4.14
7:00 PM	-119.0	3.621	6.52	1.09	176.17	29.36	238.76	4.19
8:00 PM	-118.0	3.582	6.45	1.07	174.27	29.05	236.19	4.23
9:00 PM	-117.0	3.544	6.38	1.06	172.40	28.73	233.65	4.28
10:00 PM	-116.0	3.506	6.31	1.05	170.54	28.42	231.14	4.33
11:00 PM	-115.0	3.468	6.24	1.04	168.71	28.12	228.65	4.37

SIR-Y002
5 DPC

SIR-Y002
4 DPC

SIR-Y002
3 DPC

SIR-Y002
2 DPC

SIR-Y002
1 DPC

SIR-Y002
Day of Calibration

SIR-Y002
Post-Calibration

Indication

^On average there are 27.3 million microspheres in each SIR-Y002 delivery vial (MS-TR-008).

To calculate the number of spheres drawn, divide activity drawn by the total activity in the vial and multiply by 27,300,000.000

SIR-Spheres Y-90 resin microspheres (SIR-Y002 3 mL vial)

DAY FOUR PRE-CALIBRATION

	Elapse Time (in hours)	Decay Correction Factor	SIR-Y002 Activity in GBq	SIR-Y002 Activity in GBq per .5mL	SIR-Y002 Activity in mCi	SIR-Y002 Activity in mCi per .5mL	SIR-Y002 Specific Activity Bq Per Sphere^	SIR-Y002 SIR-Spheres per GBq (reported in millions^)
12:00 AM	-114.0	3.431	6.18	1.03	166.90	27.82	226.20	4.42
1:00 AM	-113.0	3.394	6.11	1.02	165.10	27.52	223.76	4.47
2:00 AM	-112.0	3.357	6.04	1.01	163.32	27.22	221.36	4.52
3:00 AM	-111.0	3.321	5.98	1.00	161.57	26.93	218.97	4.57
4:00 AM	-110.0	3.285	5.91	0.99	159.83	26.64	216.62	4.62
5:00 AM	-109.0	3.250	5.85	0.98	158.11	26.35	214.29	4.67
6:00 AM	-108.0	3.215	5.79	0.96	156.41	26.07	211.99	4.72
7:00 AM	-107.0	3.181	5.72	0.95	154.73	25.79	209.71	4.77
8:00 AM	-106.0	3.146	5.66	0.94	153.06	25.51	207.45	4.82
9:00 AM	-105.0	3.112	5.60	0.93	151.42	25.24	205.22	4.87
10:00 AM	-104.0	3.079	5.54	0.92	149.79	24.96	203.01	4.93
11:00 AM	-103.0	3.046	5.48	0.91	148.18	24.70	200.83	4.98
12:00 PM	-102.0	3.013	5.42	0.90	146.58	24.43	198.67	5.03
1:00 PM	-101.0	2.981	5.37	0.89	145.01	24.17	196.53	5.09
2:00 PM	-100.0	2.949	5.31	0.88	143.45	23.91	194.42	5.14
3:00 PM	-99.0	2.917	5.25	0.88	141.91	23.65	192.33	5.20
4:00 PM	-98.0	2.886	5.19	0.87	140.38	23.40	190.26	5.26
5:00 PM	-97.0	2.855	5.14	0.86	138.87	23.14	188.21	5.31
6:00 PM	-96.0	2.824	5.08	0.85	137.38	22.90	186.19	5.37
7:00 PM	-95.0	2.793	5.03	0.84	135.90	22.65	184.18	5.43
8:00 PM	-94.0	2.763	4.97	0.83	134.44	22.41	182.20	5.49
9:00 PM	-93.0	2.734	4.92	0.82	132.99	22.17	180.24	5.55
10:00 PM	-92.0	2.704	4.87	0.81	131.56	21.93	178.31	5.61
11:00 PM	-91.0	2.675	4.82	0.80	130.15	21.69	176.39	5.67

SIR-Y002
5 DPC

SIR-Y002
4 DPC

SIR-Y002
3 DPC

SIR-Y002
2 DPC

SIR-Y002
1 DPC

SIR-Y002
Day of Calibration

SIR-Y002
Post-Calibration

Indication

^On average there are 27.3 million microspheres in each SIR-Y002 delivery vial (MS-TR-008).

To calculate the number of spheres drawn, divide activity drawn by the total activity in the vial and multiply by 27,300,000.000

SIR-Spheres Y-90 resin microspheres (SIR-Y002 3 mL vial)

DAY THREE PRE-CALIBRATION

	Elapse Time (in hours)	Decay Correction Factor	SIR-Y002 Activity in GBq	SIR-Y002 Activity in GBq per .5mL	SIR-Y002 Activity in mCi	SIR-Y002 Activity in mCi per .5mL	SIR-Y002 Specific Activity Bq Per Sphere^	SIR-Y002 SIR-Spheres per GBq (reported in millions^)
12:00 AM	-90.0	2.646	4.76	0.79	128.75	21.46	174.49	5.73
1:00 AM	-89.0	2.618	4.71	0.79	127.36	21.23	172.61	5.79
2:00 AM	-88.0	2.590	4.66	0.78	125.99	21.00	170.76	5.86
3:00 AM	-87.0	2.562	4.61	0.77	124.64	20.77	168.92	5.92
4:00 AM	-86.0	2.534	4.56	0.76	123.30	20.55	167.10	5.98
5:00 AM	-85.0	2.507	4.51	0.75	121.97	20.33	165.31	6.05
6:00 AM	-84.0	2.480	4.46	0.74	120.66	20.11	163.53	6.12
7:00 AM	-83.0	2.454	4.42	0.74	119.36	19.89	161.77	6.18
8:00 AM	-82.0	2.427	4.37	0.73	118.08	19.68	160.03	6.25
9:00 AM	-81.0	2.401	4.32	0.72	116.81	19.47	158.31	6.32
10:00 AM	-80.0	2.375	4.28	0.71	115.55	19.26	156.61	6.39
11:00 AM	-79.0	2.350	4.23	0.70	114.31	19.05	154.92	6.45
12:00 PM	-78.0	2.324	4.18	0.70	113.08	18.85	153.26	6.53
1:00 PM	-77.0	2.299	4.14	0.69	111.86	18.64	151.61	6.60
2:00 PM	-76.0	2.275	4.09	0.68	110.66	18.44	149.98	6.67
3:00 PM	-75.0	2.250	4.05	0.68	109.47	18.24	148.36	6.74
4:00 PM	-74.0	2.226	4.01	0.67	108.29	18.05	146.77	6.81
5:00 PM	-73.0	2.202	3.96	0.66	107.13	17.85	145.19	6.89
6:00 PM	-72.0	2.178	3.92	0.65	105.97	17.66	143.63	6.96
7:00 PM	-71.0	2.155	3.88	0.65	104.83	17.47	142.08	7.04
8:00 PM	-70.0	2.132	3.84	0.64	103.71	17.28	140.56	7.11
9:00 PM	-69.0	2.109	3.80	0.63	102.59	17.10	139.04	7.19
10:00 PM	-68.0	2.086	3.76	0.63	101.49	16.91	137.55	7.27
11:00 PM	-67.0	2.064	3.71	0.62	100.40	16.73	136.07	7.35

SIR-Y002
5 DPC

SIR-Y002
4 DPC

SIR-Y002
3 DPC

SIR-Y002
2 DPC

SIR-Y002
1 DPC

SIR-Y002
Day of Calibration

SIR-Y002
Post-Calibration

Indication

^On average there are 27.3 million microspheres in each SIR-Y002 delivery vial (MS-TR-008).

To calculate the number of spheres drawn, divide activity drawn by the total activity in the vial and multiply by 27,300,000.000

SIR-Spheres Y-90 resin microspheres (SIR-Y002 3 mL vial)

DAY TWO PRE-CALIBRATION

	Elapse Time (in hours)	Decay Correction Factor	SIR-Y002 Activity in GBq	SIR-Y002 Activity in GBq per .5mL	SIR-Y002 Activity in mCi	SIR-Y002 Activity in mCi per .5mL	SIR-Y002 Specific Activity Bq Per Sphere^	SIR-Y002 SIR-Spheres per GBq (reported in millions^)
12:00 AM	-66.0	2.042	3.67	0.61	99.32	16.55	134.61	7.43
1:00 AM	-65.0	2.020	3.64	0.61	98.25	16.37	133.16	7.51
2:00 AM	-64.0	1.998	3.60	0.60	97.19	16.20	131.73	7.59
3:00 AM	-63.0	1.976	3.56	0.59	96.15	16.02	130.31	7.67
4:00 AM	-62.0	1.955	3.52	0.59	95.11	15.85	128.91	7.76
5:00 AM	-61.0	1.934	3.48	0.58	94.09	15.68	127.52	7.84
6:00 AM	-60.0	1.913	3.44	0.57	93.08	15.51	126.15	7.93
7:00 AM	-59.0	1.893	3.41	0.57	92.08	15.35	124.79	8.01
8:00 AM	-58.0	1.872	3.37	0.56	91.09	15.18	123.45	8.10
9:00 AM	-57.0	1.852	3.33	0.56	90.11	15.02	122.12	8.19
10:00 AM	-56.0	1.832	3.30	0.55	89.14	14.86	120.81	8.28
11:00 AM	-55.0	1.813	3.26	0.54	88.18	14.70	119.51	8.37
12:00 PM	-54.0	1.793	3.23	0.54	87.23	14.54	118.22	8.46
1:00 PM	-53.0	1.774	3.19	0.53	86.29	14.38	116.95	8.55
2:00 PM	-52.0	1.755	3.16	0.53	85.36	14.23	115.70	8.64
3:00 PM	-51.0	1.736	3.12	0.52	84.45	14.07	114.45	8.74
4:00 PM	-50.0	1.717	3.09	0.52	83.54	13.92	113.22	8.83
5:00 PM	-49.0	1.699	3.06	0.51	82.64	13.77	112.00	8.93
6:00 PM	-48.0	1.680	3.02	0.50	81.75	13.63	110.80	9.03
7:00 PM	-47.0	1.662	2.99	0.50	80.87	13.48	109.61	9.12
8:00 PM	-46.0	1.644	2.96	0.49	80.00	13.33	108.43	9.22
9:00 PM	-45.0	1.627	2.93	0.49	79.14	13.19	107.26	9.32
10:00 PM	-44.0	1.609	2.90	0.48	78.29	13.05	106.11	9.42
11:00 PM	-43.0	1.592	2.87	0.48	77.45	12.91	104.97	9.53

SIR-Y002 5 DPC
SIR-Y002 4 DPC
SIR-Y002 3 DPC
SIR-Y002 2 DPC
SIR-Y002 1 DPC
SIR-Y002 Day of Calibration
SIR-Y002 Post-Calibration
Indication

^On average there are 27.3 million microspheres in each SIR-Y002 delivery vial (MS-TR-008).

To calculate the number of spheres drawn, divide activity drawn by the total activity in the vial and multiply by 27,300,000.000

SIR-Spheres Y-90 resin microspheres (SIR-Y002 3 mL vial)

DAY ONE PRE-CALIBRATION

	Elapse Time (in hours)	Decay Correction Factor	SIR-Y002 Activity in GBq	SIR-Y002 Activity in GBq per .5mL	SIR-Y002 Activity in mCi	SIR-Y002 Activity in mCi per .5mL	SIR-Y002 Specific Activity Bq Per Sphere^	SIR-Y002 SIR-Spheres per GBq (reported in millions^)
12:00 AM	-42.0	1.575	2.83	0.47	76.61	12.77	103.84	9.63
1:00 AM	-41.0	1.558	2.80	0.47	75.79	12.63	102.72	9.74
2:00 AM	-40.0	1.541	2.77	0.46	74.98	12.50	101.62	9.84
3:00 AM	-39.0	1.525	2.74	0.46	74.17	12.36	100.52	9.95
4:00 AM	-38.0	1.508	2.71	0.45	73.37	12.23	99.44	10.06
5:00 AM	-37.0	1.492	2.69	0.45	72.58	12.10	98.37	10.17
6:00 AM	-36.0	1.476	2.66	0.44	71.80	11.97	97.31	10.28
7:00 AM	-35.0	1.460	2.63	0.44	71.03	11.84	96.27	10.39
8:00 AM	-34.0	1.444	2.60	0.43	70.27	11.71	95.23	10.50
9:00 AM	-33.0	1.429	2.57	0.43	69.51	11.58	94.21	10.61
10:00 AM	-32.0	1.413	2.54	0.42	68.76	11.46	93.19	10.73
11:00 AM	-31.0	1.398	2.52	0.42	68.02	11.34	92.19	10.85
12:00 PM	-30.0	1.383	2.49	0.41	67.29	11.22	91.20	10.96
1:00 PM	-29.0	1.368	2.46	0.41	66.57	11.09	90.22	11.08
2:00 PM	-28.0	1.354	2.44	0.41	65.85	10.98	89.25	11.20
3:00 PM	-27.0	1.339	2.41	0.40	65.14	10.86	88.29	11.33
4:00 PM	-26.0	1.325	2.38	0.40	64.44	10.74	87.34	11.45
5:00 PM	-25.0	1.310	2.36	0.39	63.75	10.62	86.40	11.57
6:00 PM	-24.0	1.296	2.33	0.39	63.06	10.51	85.47	11.70
7:00 PM	-23.0	1.282	2.31	0.38	62.39	10.40	84.55	11.83
8:00 PM	-22.0	1.269	2.28	0.38	61.71	10.29	83.64	11.96
9:00 PM	-21.0	1.255	2.26	0.38	61.05	10.18	82.74	12.09
10:00 PM	-20.0	1.241	2.23	0.37	60.39	10.07	81.85	12.22
11:00 PM	-19.0	1.228	2.21	0.37	59.74	9.96	80.97	12.35

SIR-Y002 5 DPC
SIR-Y002 4 DPC
SIR-Y002 3 DPC
SIR-Y002 2 DPC
SIR-Y002 1 DPC
SIR-Y002 Day of Calibration
SIR-Y002 Post-Calibration
Indication

^On average there are 27.3 million microspheres in each SIR-Y002 delivery vial (MS-TR-008).

To calculate the number of spheres drawn, divide activity drawn by the total activity in the vial and multiply by 27,300,000.000

SIR-Spheres Y-90 resin microspheres (SIR-Y002 3 mL vial)

DAY OF CALIBRATION

	Elapse Time (in hours)	Decay Correction Factor	SIR-Y002 Activity in GBq	SIR-Y002 Activity in GBq per .5mL	SIR-Y002 Activity in mCi	SIR-Y002 Activity in mCi per .5mL	SIR-Y002 Specific Activity Bq Per Sphere^	SIR-Y002 SIR-Spheres per GBq (reported in millions^)
12:00 AM	-18.0	1.215	2.19	0.36	59.10	9.85	80.10	12.48
1:00 AM	-17.0	1.202	2.16	0.36	58.47	9.74	79.24	12.62
2:00 AM	-16.0	1.189	2.14	0.36	57.84	9.64	78.39	12.76
3:00 AM	-15.0	1.176	2.12	0.35	57.22	9.54	77.54	12.90
4:00 AM	-14.0	1.163	2.09	0.35	56.60	9.43	76.71	13.04
5:00 AM	-13.0	1.151	2.07	0.35	55.99	9.33	75.89	13.18
6:00 AM	-12.0	1.139	2.05	0.34	55.39	9.23	75.07	13.32
7:00 AM	-11.0	1.126	2.03	0.34	54.79	9.13	74.26	13.47
8:00 AM	-10.0	1.114	2.01	0.33	54.20	9.03	73.46	13.61
9:00 AM	-9.0	1.102	1.98	0.33	53.62	8.94	72.67	13.76
10:00 AM	-8.0	1.090	1.96	0.33	53.04	8.84	71.89	13.91
11:00 AM	-7.0	1.079	1.94	0.32	52.47	8.75	71.12	14.06
12:00 PM	-6.0	1.067	1.92	0.32	51.91	8.65	70.35	14.21
1:00 PM	-5.0	1.056	1.90	0.32	51.35	8.56	69.60	14.37
2:00 PM	-4.0	1.044	1.88	0.31	50.80	8.47	68.85	14.52
3:00 PM	-3.0	1.033	1.86	0.31	50.25	8.38	68.11	14.68
4:00 PM	-2.0	1.022	1.84	0.31	49.71	8.29	67.38	14.84
5:00 PM	-1.0	1.011	1.82	0.30	49.18	8.20	66.65	15.00
6:00 PM	0.0	1.000	1.80	0.30	48.65	8.11	65.93	15.17
7:00 PM	1.0	0.989	1.78	0.30	48.13	8.02	65.22	15.33
8:00 PM	2.0	0.979	1.76	0.29	47.61	7.93	64.52	15.50
9:00 PM	3.0	0.968	1.74	0.29	47.10	7.85	63.83	15.67
10:00 PM	4.0	0.958	1.72	0.29	46.59	7.76	63.14	15.84
11:00 PM	5.0	0.947	1.71	0.28	46.09	7.68	62.46	16.01

SIR-Y002
5 DPC

SIR-Y002
4 DPC

SIR-Y002
3 DPC

SIR-Y002
2 DPC

SIR-Y002
1 DPC

SIR-Y002
Day of Calibration

SIR-Y002
Post-Calibration

Indication

^On average there are 27.3 million microspheres in each SIR-Y002 delivery vial (MS-TR-008).

To calculate the number of spheres drawn, divide activity drawn by the total activity in the vial and multiply by 27,300,000.000

SIR-Spheres Y-90 resin microspheres (SIR-Y002 3 mL vial)

POST-CALIBRATION

	Elapse Time (in hours)	Decay Correction Factor	SIR-Y002 Activity in GBq	SIR-Y002 Activity in GBq per .5mL	SIR-Y002 Activity in mCi	SIR-Y002 Activity in mCi per .5mL	SIR-Y002 Specific Activity Bq Per Sphere^	SIR-Y002 SIR-Spheres per GBq (reported in millions^)
12:00 AM	6.0	0.937	1.69	0.28	45.59	7.60	61.79	16.18
1:00 AM	7.0	0.927	1.67	0.28	45.10	7.52	61.13	16.36
2:00 AM	8.0	0.917	1.65	0.28	44.62	7.44	60.47	16.54
3:00 AM	9.0	0.907	1.63	0.27	44.14	7.36	59.82	16.72
4:00 AM	10.0	0.898	1.62	0.27	43.66	7.28	59.18	16.90
5:00 AM	11.0	0.888	1.60	0.27	43.19	7.20	58.54	17.08
6:00 AM	12.0	0.878	1.58	0.26	42.73	7.12	57.91	17.27
7:00 AM	13.0	0.869	1.56	0.26	42.27	7.04	57.29	17.46
8:00 AM	14.0	0.860	1.55	0.26	41.81	6.97	56.67	17.65
9:00 AM	15.0	0.850	1.53	0.26	41.36	6.89	56.06	17.84
10:00 AM	16.0	0.841	1.51	0.25	40.92	6.82	55.46	18.03
11:00 AM	17.0	0.832	1.50	0.25	40.48	6.75	54.86	18.23
12:00 PM	18.0	0.823	1.48	0.25	40.04	6.67	54.27	18.43
1:00 PM	19.0	0.814	1.47	0.24	39.61	6.60	53.69	18.63
2:00 PM	20.0	0.806	1.45	0.24	39.19	6.53	53.11	18.83
3:00 PM	21.0	0.797	1.43	0.24	38.77	6.46	52.54	19.03
4:00 PM	22.0	0.788	1.42	0.24	38.35	6.39	51.97	19.24
5:00 PM	23.0	0.780	1.40	0.23	37.94	6.32	51.42	19.45
6:00 PM	24.0	0.771	1.39	0.23	37.53	6.25	50.86	19.66

SIR-Y002
5 DPC

SIR-Y002
4 DPC

SIR-Y002
3 DPC

SIR-Y002
2 DPC

SIR-Y002
1 DPC

SIR-Y002
Day of Calibration

SIR-Y002
Post-Calibration

Indication

^On average there are 27.3 million microspheres in each SIR-Y002 delivery vial (MS-TR-008).

To calculate the number of spheres drawn, divide activity drawn by the total activity in the vial and multiply by 27,300,000.000

Caution: Federal (USA) law restricts this device to sale by or on the order of a physician.

Indications for Use: SIR-Spheres Y-90 resin microspheres are indicated for the treatment of unresectable metastatic liver tumors from primary colorectal cancer with adjuvant intrahepatic artery chemotherapy (IHAC) of FUDR (Floxuridine). **Warnings / Precautions:** Inadvertent delivery of the microspheres to locations other than the intended hepatic tumor may result in local radiation damage. Due to the radioactivity and the significant consequences of misplacing the microspheres in situ, this product must be implanted by physicians who have completed the Sirtex TEC training program. A SPECT scan of the upper abdomen immediately after implantation is recommended. Patients may experience abdominal pain immediately after administration and pain relief may be required. H-2 blocking agents may be administered the day before implantation and continued as needed to reduce gastric complications. **Side Effects:** Common side effects are fever, transient decrease of hemoglobin, mild to moderate abnormality of liver function tests, abdominal pain, nausea, vomiting, and diarrhea. Potential serious effects due to exposure to high radiation include acute pancreatitis, radiation pneumonitis, acute gastritis, radiation hepatitis, and acute cholecystitis. **Contraindications:** SIR-Spheres Y-90 resin microspheres should not be implanted in patients who have either had previous external beam radiation therapy to the liver; ascites or are in clinical liver failure. This device is contraindicated in patients with markedly abnormal synthetic and excretory liver function tests; greater than 20% lung shunting of the hepatic artery blood flow, or >30 Gy radiation absorbed dose to the lungs, as determined by the ^{99m}Tc MAA scan; disseminated extra-hepatic malignant disease, and portal vein thrombosis. This device should not be implanted in patients determined via angiogram to have an abnormal vascular anatomy that would result in significant reflux of the hepatic arterial blood flow to the stomach, pancreas, or bowel. **General Information:** SIR-Spheres Y-90 resin microspheres may only be distributed to a duly licensed or accredited facility capable of handling therapeutic medical isotopes. This product is radioactive and should thus be handled in accordance with all applicable standards and regulations. **Consult the Instructions for Use (www.sirtex.com) for a complete listing of indications, contraindications, side effects, warnings, and precautions.**

SIR-Y002
5 DPC

SIR-Y002
4 DPC

SIR-Y002
3 DPC

SIR-Y002
2 DPC

SIR-Y002
1 DPC

SIR-Y002
Day of Calibration

SIR-Y002
Post-Calibration

Indication



Manufacturer

Sirtex Medical Pty Ltd
Shop 6, 207 Pacific Highway
St Leonards, NSW 2065
Australia
Tel: +61 2 9964 8400
Fax: +61 2 9964 8410

Americas

Sirtex Medical Inc.
300 Unicorn Park Drive
Woburn, MA 01801
USA
Tel: +1 888 474 7839

www.sirtex.com

SIR-Spheres® is a registered trademark of Sirtex SIR-Spheres Pty Ltd.
©2024 Sirtex Medical Inc. APM-US-003-07-24 V4

SIRTeX